



NC DEPARTMENT OF  
**HEALTH AND  
HUMAN SERVICES**  
Division of Public Health

**ROY COOPER** • Governor

**MANDY COHEN, MD, MPH** • Secretary

**MARK T. BENTON** • Assistant Secretary for Public Health

Division of Public Health

**SCOTT M. SHONE, PhD, HCLD (ABB)** • Laboratory Director, State  
Laboratory of Public Health

## Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Martha Richardson, MLT (ASCP)  
Medical Laboratory Technologist

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** November 09, 2020

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Bio-Bag™ Type C:** Lot Number: **B02C006M** Expiration Date: **05/31/2021**

### Summary of NCSLPH QC results for this Lot #:

*N. gonorrhoeae* ATCC# 43069 - Growth

**Lot number failed:** None  
Hardy Diagnostics™ Modified Thayer Martin media

\*\*If your contact information changes, please provide that information to Martha Richardson in Laboratory Improvement at (919) 733-7186 or via e-mail at [martha.richardson@dhhs.nc.gov](mailto:martha.richardson@dhhs.nc.gov). Thank you.