



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

ROY COOPER • Governor
MANDY COHEN, MD, MPH • Secretary
MARK T. BENTON • Assistant Secretary for Public Health
Division of Public Health
SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Martha Richardson, MLT (ASCP)
Laboratory Improvement Consultant

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: September 21, 2020

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: **0199595** Expiration Date: **11/10/2020**
Along with **Bio-Bag™ Type C:** Lot Number: **B02B133M** Expiration Date: **10/31/2020**
and with **Bio-Bag™ Type C:** Lot Number: **B02C099M** Expiration Date: **10/30/2021**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
<i>N. gonorrhoeae</i> ATCC# 43070 - Growth	<i>P. mirabilis</i> ATCC# 43071 – Partial Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
BD and Fisher Scientific contacted: September 18, 2020

**If your contact information changes, please provide that information to Martha Richardson in Laboratory Improvement at (919) 733-7186 or via e-mail at martha.richardson@dhhs.nc.gov. Thank you.