



NC DEPARTMENT OF  
**HEALTH AND  
HUMAN SERVICES**  
Division of Public Health

**ROY COOPER** • Governor

**MANDY COHEN, MD, MPH** • Secretary

**MARK T. BENTON** • Assistant Secretary for Public Health

Division of Public Health

**SCOTT M. SHONE, PhD, HCLD (ABB)** • Laboratory Director, State  
Laboratory of Public Health

## Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Martha Richardson, MLT (ASCP)  
Laboratory Improvement Consultant

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** November 5, 2020

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Modified Thayer Martin II:** Lot Number: 472436 Expiration Date: **01/05/2021**  
Along with **Bio-Bag™ Type C:** Lot Number: B02C006M Expiration Date: **05/31/2021**  
and with **Bio-Bag™ Type C:** Lot Number: B02C099M Expiration Date: **10/30/2021**

### Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 –No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None  
Hardy Diagnostics Modified Thayer Martin

\*\*If your contact information changes, please provide that information to Martha Richardson in Laboratory Improvement at (919) 733-7186 or via e-mail at martha.richardson@dhhs.nc.gov. Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

www.ncdhhs.gov • <http://publichealth.nc.gov> • [slph.ncpublichealth.com](http://slph.ncpublichealth.com) • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER