



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

ROY COOPER • Governor

MANDY COHEN, MD, MPH • Secretary

MARK T. BENTON • Assistant Secretary for Public Health

Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State
Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT (ASCP)
Laboratory Improvement Medical Laboratory Technologist

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: January 14th, 2021

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Modified Thayer Martin II: Lot Number: **475566** Expiration Date: **02/23/2021**
Along with **Bio-Bag™ Type C:** Lot Number: **B02C006M** Expiration Date: **05/31/2021**
and with **Bio-Bag™ Type C:** Lot Number: **B02C099M** Expiration Date: **10/30/2021**

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 –No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

Lot number failed: None
Hardy Diagnostics Modified Thayer Martin

**If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

www.ncdhhs.gov • <http://publichealth.nc.gov> • slph.ncpublichealth.com • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER