



NC DEPARTMENT OF  
**HEALTH AND  
HUMAN SERVICES**  
Division of Public Health

ROY COOPER • Governor

MANDY COHEN, MD, MPH • Secretary

MARK T. BENTON • Assistant Secretary for Public Health

Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State  
Laboratory of Public Health

## Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Angela Bradley, MLT (ASCP)  
Laboratory Improvement Medical Laboratory Technologist

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** February 11th, 2021

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Modified Thayer Martin II:** Lot Number: **477621** Expiration Date: **03/30/2021**  
Along with **Bio-Bag™ Type C:** Lot Number: **B02C006M** Expiration Date: **05/31/2021**  
and with **Bio-Bag™ Type C:** Lot Number: **B02C099M** Expiration Date: **10/30/2021**

### Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
	<i>P. mirabilis</i> ATCC# 43071 –No Growth
Sterility Check - Acceptable	Visual Check - Acceptable

**Lot number failed:** None  
Hardy Diagnostics Modified Thayer Martin

\*\*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at [angela.bradley@dhhs.nc.gov](mailto:angela.bradley@dhhs.nc.gov). Thank you.